

**Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 1/26/2015**

lms
mmb

Instrument Serial Number: 012222

As of 26-Jan-15

Start Date	End Date	Agency Name
22-Jan-15		DFS Central Lab
06-Aug-14	22-Jan-15	Lee County Sheriff's Office
29-Jul-14	06-Aug-14	DFS Central Lab
26-Nov-13	29-Jul-14	Chesterfield County PD
10-Dec-12	26-Nov-13	DFS Central Lab
26-Oct-12	10-Dec-12	Chesapeake PD
16-May-12	26-Oct-12	DFS Central Lab

JMS
mmB

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 012222 Worksheet Start Date 1/22/2015

Location Lee County

Address 33640 Main St., Jonesville, VA 24263

DFS Technician Heather Stanton License No. 18910

☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 725 Reference Barometer (mm HG) 731

Reference Barometer(RB)Serial # 009113 RB Calibration Due 4/28/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.294
0.300	0.286	0.277	0.294	Sample 2	0.294
Precision		sample min	sample max	Sample 3	0.293
0.001		0.293	0.294		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.097
0.100	0.095	0.092	0.098	Sample 2	0.097
Precision		sample min	sample max	Sample 3	0.097
0		0.097	0.097		

Dry gas standard Lot No. (with tank no.) AG400202-31

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Instrument Serial Number

012222

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath M. Stanton

Date

1/26/15

Issuing Analyst


Michelle Buller

Date

1/27/15

JPD

AMS
mmB




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 01/22/2015
TEST EQUIPMENT NUMBER 012222		

RESULTS: TIME SAMPLE TAKEN **09:16 EST**

SAMPLE'S ALCOHOL CONTENT **0.00** GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 22 Jan 2015 at 10:11

lms
mmb

Test Results

Instrument Serial Number 012222

Test # 000960 Subject Test

Test Location 1 Department of
Test Date 22 Jan 2015

Test Location 2 Forensic Science
Test Time 09:10
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

Effective Date 10/01/2014

License Number 18910

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 22 Jan 2015 End Time 09:17

Result Time 09:16

Result Date 22 Jan 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 09:11

Data Type BLK

Sample Value 0.000

Sample Time 09:11

Data Type CHK

Sample Value 0.097

Sample Time 09:12

Data Type BLK

Sample Value 0.000

Sample Time 09:13

Data Type SUBJ

Sample Value 0.000

Sample Time 09:14

Data Type BLK

Sample Value 0.000

Sample Time 09:15

Data Type SUBJ

Sample Value 0.000

Sample Time 09:16

Data Type BLK

Sample Value 0.000

Sample Time 09:17

Standard Type Dry Gas Std

Standard Value 0.095

Standard Lot Number AG400202-31

Standard Expiration Date 01/02/2016

Tank Pressure 919

Barometric Pressure 725 mmHg

Blow Sample Number 1 Blow Duration 3.20 sec

Blow Volume 1747 cc End-of-Blow Time 09:14

Blow Sample Number 2 Blow Duration 3.43 sec

Blow Volume 1775 cc End-of-Blow Time 09:16

Tamper Evident Stamp b1b0d84f

Test Status Code 0

Test Status Success

HS
mmB

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 012222 Test Number: 961

Test Date: 01/22/2015 Test Time: 09:18 EST

Dry Gas Target: 0.095

Lot Number: AG400202-31 Exp Date: 01/02/2016

Tank Pressure: 914 psi Barometric Pressure: 725 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	09:19
CHK	0.097	09:20
BLK	0.000	09:21
CHK	0.097	09:22
BLK	0.000	09:23
CHK	0.097	09:24

Test Status: *Success*

Calibration CRC: 20E3A79E

1110

mmB

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 012222 Test Number: 962

Test Date: 01/22/2015 Test Time: 09:28 EST

Dry Gas Target: 0.286

Lot Number: AG428003-09 Exp Date: 10/07/2016

Tank Pressure: 272 psi Barometric Pressure: 725 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	09:28
CHK	0.294	09:29
BLK	0.000	09:31
CHK	0.294	09:32
BLK	0.000	09:33
CHK	0.293	09:34

Test Status: *Success*

Calibration CRC: 20E3A79E